

IMPROVED AQUA REGIA METAL EXTRACTION USING A LOAM SOIL CRM ERC-CC141 AND A PERFORMANT MICROWAVE OVEN

Nicoleta Mirela Marin^{1,2}, Gabriela Geanina Vasile¹, Jana Petre¹,
Liliana Valeria Cruceru¹

¹ National Research and Development Institute for Industrial Ecology – Bucharest, 71-73 Podul Dambovitei Street, 060650, Bucharest, Romania
e-mail: nicoleta.marin@incdecoind.ro

² Department of Analytical Chemistry, Faculty of Chemistry, University of Bucharest, 2-14 Blvd. Regina Elisabeta, 030018, Bucharest, Romania,

Abstract

In this study a digestion method for determination of metals (Co, Cr, Cu, Ni, Pb, Zn) in soil aqua regia extracts was performed using a microwave digestion technique combined with ICP-EOS and FAAS techniques. Improved digestion program was applied in three steps at 220°C maximum temperature for 35 minutes to 1 g of soil with a mixture of HCl and HNO₃ (report 6 mL to 2 mL). In experimental studies was used a Certified Reference Material CRM ERC-CC141 loam soil. The recovery percentages in CRM standard were situated for all metals in the range 93.5% to 105% for ICP-EOS determinations and 92% to 107% for FAAS determinations.

Keywords: soil, metals, microwave digestion methods, ICP-EOS, FAAS

Introduction

Normally in the soil there are small amounts of metals, but sometimes their levels increased due to the anthropic sources and industrial pollution. The soil behaves as a deposit for the metallic elements due to its high accumulation capacity during the time. Based on the above observations, the monitoring of the metals in this type of matrix is important [1-5].

Accurate determinations of metals in soil samples lead to development and / or improvement of digestion methods as well as choosing appropriate technique for obtaining reproducible results. Microwave assisted digestion provides a fast and suitable method for soil complex matrices [5-9].

The objectives of the present study are: (i) optimization of a microwave digestion method for determination of metals in CRM soil using Flame Atomic Absorption Spectrometry (FAAS) and Inductively Coupled Plasma with Optical Emission Spectrometry (ICP-EOS) techniques. (ii) extending the application of this method for agricultural and industrial soil samples.

2. Materials and Methods

2.1 Equipments

FAAS Thermo Scientific M6 Dual Spectrometer and the ICP-OES Optima 5300 DV Perkin Elmer Spectrometer were used for determination of Co, Cr, Cu, Ni, Pb and Zn concentrations. The digestion of samples was done with a Microwave Digestion System Ethos Up Milestone.

2.2 Reagents and calibration

The acids used, hydrochloric acid 37% and nitric acids 67%, were reagent quality from Merk.

Ultra-purity water was produced with a Millipore Milli-Q System.

For the calibration curves were prepared standards in the range 0.5 to 2.5 mg/L using a 100 mg/L Multielement Certified Reference Material (ICP multielement standard solution XVI, Certipur, Merck) for ICP-EOS technique. For FAAS determination were used unielement standards for AAS, the calibration curves were plotted in the range 1 to 8 mg/L. For the matrix solution, the standards were prepared with aqua regia in the same proportion as was used for digestion step. Quality control of the analytical results was performed with a multielement Certified Reference Material for ICP, 1000 mg/L, Merck quality.

FAAS analytical determinations using an air-acetylene flame were carried out based on the wavelengths presented in Table 1. ICP-EOS parameters were presented in Table 2.

Table 1. The adsorption and emission lines used in FAAS and ICP-OES

Metal	FAAS Wavelength (nm)	ICP-OES Wavelength (nm)
Pb	217.0	220.353
Co	240.7	228.616
Cr	357.9	267.716
Cu	324.8	327.393
Ni	232.0	231.604
Zn	213.9	206.200

Table 2. Spectrometer operating parameters

<i>ICP-OES Spectrometer parameters</i>			
Delay time:	40 s	Purge gas flow:	normal
Replicates	2 times	Peristaltic pump:	1.5 mL/min
<i>Plasma parameters</i>			
Plasma flow rate:	15 L/min	Power RF:	1400 W
Auxiliar flow rate:	0.2 L/min	Plasma view:	axial
Nebulizer flow rate:	0.7 L/min	View distance:	15.0 mm
<i>Spectral peak processing</i>			
Peak algorithm:	Peak area	Points per peak:	7 points

2.3 Aqua regia extraction

Over 1 g of CRM sample was added 2 ml HNO₃ and 6 mL HCl, samples were weighed into microwave oven vessels and placed in the microwave oven. In order to obtain suitable extracts was performed a multi-steps digestion program: for 15 minutes the temperature inside the vessels was increased to 220°C, then was kept for 20 minutes at 220°C and the cooling process was applied for 30 minutes. The steps of the digestion program is given in table 3. After cooled, the solutions were filtered on 0,45 µm porosity paper in 50 mL volumetric flasks and were analysed with FAAS and ICP-EOS techniques. Temperature T₂ is controlled outside of the reaction vessels by an IR sensor.

Table 3. Digestion program

Step	T ₁ °C	T ₂ °C	Power (W)	Time (min)
1	220	120	1800	15
2	220	120	1800	20
3	cooled	cooled	-	30

3. Results and Discussion

In order to evaluate the extraction procedure, the obtained experimental values were compared with certified results of the CRM ERC-CC141 loam soil. In table 4 and 5 are presented comparative results for both techniques applied. The recovery percentages obtained suggests that the method is adequate for intended purpose.

Table 4. Determinations of metals in loam soil using ICP

Metallic elements	CRM ERC-CC141 loam soil		
	Obtained value mg/kg d.m.*	Certified value mg/kg d.m.*	Recovery (%)
Pb	30.1 ± 3.6	32.2 ± 4.0	93.5
Co	8.20 ± 1.0	7.90 ± 0.90	104
Cr	30.5 ± 5.3	31.0 ± 4.0	98.0
Cu	12.5 ± 1.4	12.4 ± 0.9	101
Ni	21.4 ± 2.3	21.9 ± 1.6	98.0
Zn	51.1 ± 5.4	50.0 ± 4.0	102

*dry matter

Table 5. Determinations of metals in loam soil using FAAS

Metallic elements	CRM ERC-CC141 loam soil		
	Obtained value mg/kg d.m.*	Certified value mg/kg d.m.*	Recovery (%)
Pb	29.6 ± 3.5	32.2 ± 4.0	92
Co	7.70 ± 0.96	7.90 ± 0.90	97
Cr	31.6 ± 4.9	31.0 ± 4.0	102
Cu	11.6 ± 1.3	12.4 ± 0.9	93
Ni	21.6 ± 2.0	21.9 ± 1.6	99
Zn	53.4 ± 5.0	50.0 ± 4.0	107

*dry matter

The improved digestion method was applied to eight soil samples collected from 0 to 5 cm depth from an industrial area situated in Ilfov region. The samples were analyzed both with FAAS and ICP-EOS techniques and the results for the metals content are presented in table 6 and 7. Both techniques provide similar results in the range of associated measurement uncertainty. The values of Co, Cr and Zn are within the normal values according with Romanian legislation for soil quality [10]. The results obtained for Pb, Ni and Cu are over normal values, but were below the alert threshold for industrial samples.

Table 6. Determinations of metals in industrial soil by ICP-OES

Metallic elements	Unit	Normal Values	S1	S2	S3	S4	S5	S6	S7
Pb	mg/kg d.m.	20	26.7	23.3	14.0	9.04	10.7	16.0	13.6
Co	mg/kg d.m.	15	7.43	8.77	9.88	7.71	8.84	7.81	9.26
Cr	mg/kg d.m.	30	22.7	28.2	35.2	18.5	31.4	14.4	21.0
Cu	mg/kg d.m.	20	39.5	36.4	45.7	23.0	40.3	20.2	23.1
Ni	mg/kg d.m.	20	27.7	43.2	41.7	26.5	32.8	24.4	29.1
Zn	mg/kg d.m.	100	99.1	97.5	65.6	40.2	51.5	54.3	53.8

Table 7. Determinations of metals in industrial soil by FAAS

Metallic elements	Unit	Normal Values	S1	S2	S3	S4	S5	S6	S7
Pb	mg/kg d.m.	20	27.7	24.3	14.9	9.20	10.9	16.9	14.1
Co	mg/kg d.m.	15	7.73	8.99	10.2	8.23	9.31	8.12	10.3
Cr	mg/kg d.m.	30	23.9	29.3	35.9	19.6	32.2	14.9	22.3
Cu	mg/kg d.m.	20	40.5	37.4	46.5	24.1	39.9	20.9	23.7
Ni	mg/kg d.m.	20	29.7	44.2	42.7	27.1	33.1	25.2	30.6
Zn	mg/kg d.m.	100	99.0	99.5	70.6	43.2	53.5	55.3	54.5

4. Conclusions

The study proposes an improved microwave aqua regia extraction method for metals determination. The results obtained after digestion of two subsequent samples from CRM ERC-CC141 loam soil shows very good recovery percentages (92% - 107%) for metal Co, Cr, Cu, Ni, Pb, Zn concentrations, results obtained with ICP-EOS and FAAS techniques.

5. References

- [1] Ravanbakhsh S, Zohreh M T, Fereshteh D, (2008), Microwave assisted digestion of soil, sludge and sediment for determination of heavy metals with ICP-OES and FAAS, *Rasayan J. Chem.*, **1**, 757-765.
- [2] Xue Z J, Liu S. Q, Liu Y L, Yan Y L, (2012), Health Risk Assessment of Heavy Metals for Edible Parts of Vegetables Grown in Sewage-Irrigated Soils in Suburbs of Baoding City, China, *Environmental Monitoring and Assessment*, **184(6)**, 3503-3513.
- [3] Wang M, Markert B, Chen W, Peng C, Ouyang Z J, (2012), Identification of Heavy Metal Pollutants Using Multivariate Analysis and Effects of Land Uses on Their Accumulation in Urban Soils in Beijing, China, *Environmental Monitoring and Assessment*, **184(10)**, 5889-5897.

- [4] Micó C, Peris M, Sánchez J, Recatalá L, (2006), Heavy metal content of agricultural soils in a Mediterranean semiarid area: the Segura River Valley (Alicante, Spain), *Spanish Journal of Agricultural Research*, **4**(4), 363-372.
- [5] Hernández-Mendoza H, Mejuto M, Cardona A. I, García-Álvarez A, Rocio Millán R, Yllera A, (2013), Optimization and Validation of a Method for Heavy Metals Quantification in Soil Samples by Inductively Coupled Plasma Sector Field Mass Spectrometry (ICP-SFMS), *American Journal of Analytical Chemistry*, **4**, 9-15.
- [6] Bettinelli M, Beone G M, Spezia S, Baffi C, (2000), Determination of heavy metals in soils and sediments by microwave-assisted digestion and inductively coupled plasma optical emission spectrometry analysis, **424**, 289-296.
- [7] Kingston H M, Haswell S J, (1997), *Microwave-Enhanced Chemistry, Fundamentals, Sample Preparation and Applications*, American Chemical Society, Washington DC.
- [8] Thulasya Ramanathan T, Yen-Peng T,(2015), Selection of wet digestion methods for metal quantification in hazardous solid wastes, *Journal of Environmental Chemical Engineering*, **3**, 1459–1467.
- [9] Samuel Melaku S, Dams R, Moens L, (2005), Determination of trace elements in agricultural soil samples by inductively coupled plasma-mass spectrometry: Microwave acid digestion versus aqua regia extraction, *Analytica Chimica Acta*, **543**, 117–123.
- [10] Order 756/1997 – Reference Values for trace chemical elements in soil, Romanian Official Gazette part 1, no. 303 bis, 08.11.1997.